

## High reliability LUX METER series

Complies with **DIN Class B** and **JIS Class AA** Compatible with **LED/OLED** lighting



Built-in Bluetooth® wireless technology FT3425

From measurement to report creation

Cut work time in half










# Ideal for low-illuminance measurement Support for measurement of 1 lx

20 lx range measurement resolution  $0.01 \ lx$ 

### Large, easy-to-see LCD display

The backlight turns on automatically whenever a measured value is retained in a low-illuminance environment.

## Measure with sensor and display units undocked



Sensor unit and main display can be separated to 2m, letting you measure at a distance away from the sensor in order to accommodate for difficult locations, shadows, and other issues.

CONNECTION CABLE L9820 (Option)

#### **Timer hold function**

Retain the measured value after a user-selected amount of time has elapsed from the time the TIMER key is pressed. In this way, you can time measurement to occur after you have moved away from the lux meter so that measurement is not affected by clothing, shadows, etc.



Select from 5 / 10 / 15 / 20 / 30 / 45 / 60 sec.

#### Remaining time display

Counts down with timer.

#### After the set time has elapsed

The measured value is retained.  $\rightarrow$  The backlight turns on and the beep sounds for 3 sec.







Measure without needing to crouch close to the ground. Also convenient for repeated measurements.

## Recommended points

### Reduce your physical burden EXTENSION CART Z5023 (option)

Hioki offers an auxiliary cart equipped with caster wheels so that it can be easily moved between measurement locations. The cart makes the measurement process significantly less physically demanding by eliminating the need to squat down to position the instrument or read its display. When using the FT3425 with a smartphone or tablet, there's no need for a connection cable (see photograph on the first page of this catalog).

## Key Features

### Memory function makes multipoint measurement a breeze

#### Memory function (up to 99 values)

Save measured values for multiple measurement locations in the instrument's internal memory on the spot for later display at your convenience.

#### Data communications functionality

Transfer data saved in the instrument's internal memory to a computer via a USB connection. Data can be saved as a text file.

#### Other software functionalities

- Display graphs and save files for user-specified time intervals. (Data can also be saved manually.)
- Display measured values on a computer screen in real time.



### Record variations in illuminance with D/A output.

Output cord must be modified to suit the connected device. (Use a commercially available USB power adapter to supply power for extended periods of time.)



## Cut work time in half!

FT3425 Built-in Bluetooth® wireless technology

Shorter work times

No more errors



### Multi-point measurement capability is ideal for final inspections of electrical and lighting work

Inspect (Illuminance measurement)

Record results

- Measure and record results in all rooms where work was performed.
- ✓ The number of measurement locations ranges from a few points in single rooms to tens of thousands of points on large floors.
- Measurements must be made after the building is complete but before furnishings are installed, resulting in a rushed schedule and sometimes requiring work to be performed at night.
- ✓ Workers must compare readings with design data.

"We want to complete the process of inspecting and recording results for numerous locations quickly!"

ticular locations on drawings shown on the phone's display


Create report

Present to client

- Create a report based on the recorded measured values after returning to the office (where mistakes are likely due to reliance on visual observations, handwritten notes, and copying of results).
- ✓ Submit the report to the client

"We want to accurately summarize an enormous volume of recorded data in a report!"





measured values, jotting down handwritten memos, and entering data!

#### FT3424, FT3425 Specifications (Accuracy guaranteed for 2 years, Post-adjustment accuracy guaranteed for 2 years)

Only FT3425 is equipped with Bluetooth® wireless technology, others are shared specifications					
Classification	DIN 5032-7: 1985 class B JIS C 1609-1: 2006 general AA class				
Light receiving element	Silicon photo-diode				
Display	Display: 4 digit, 2000 count LCD Display unit: Ix (lux) Display update rate: 500 ms ±20 ms				
	Range	Measurement ra	ange	Display step	
	20 lx	0.00 lx to 20.	.00 lx	1 count/step	
Measurement ranges	200 lx	0.0 lx to 200	0.0 <i>lx</i>		
	2000 lx	0 lx to 20	000 lx		
	20000 lx	00 lx to 200	000 lx	10 counts/step	
	200000 lx	000 lx to 2000	000 lx	100 counts/step	
Range selection	Auto/Manual				
Linearity	±2% rdg. (Multiply by 1.5 for display values in excess of 3000 lx.)				
Accuracy guarantee conditions	Sensor unit and display unit must bear the same identification number.				
Accuracy guarantee for					
temperature and humidity	21°C to 27°C (69.8°F to 80.6°F), 75% RH or less (non-condensing)				
Characteristics	[Temperature characteristics] ±3% rdg. [Humidity characteristics] ±3% rdg.				
Response time	Auto range: within 5 seconds, Manual range: within 2 seconds				
Output specifications	_		Range	Output rate	
Output method	: D/A output		20 lx	1 mV DC/ 0.01 lx	
	: 2 V/range f.s.		200 lx	1 mV DC/ 0.1 lx	
Resolution			2000 lx	1 mV DC/ 1 lx	
Output update rate		S	20000 lx	1 mV DC/ 10 lx	
Output resistance		V(at output rate)	200000 lx	1 mV DC/ 100 lx	
Power supply	: ±1% rdg. ±5 mV(at output rate) 20000 tx 1 my DC/100 tx  AA/LR6 alkaline battery ×2, R6 Manganese battery ×2, USB bus power 5 V DC				
1 Ower supply	Approx. 300 hours				
Continuous battery		a alkaline batteries, no Bluetooth® wireless technology) urs			
operation time	Approx. 80 hour				
	(when using AA alkaline batteries, with Bluetooth® wireless technology				
Auto-power off	Turns off the instrument 10 min. ±1 min. after the last key operation (can be canceled).				
Operating temperature and humidity	-10°C to 40°C (14°F to 104°F), 80% RH or less (non-condensing)				
Storage temperature and humidity	-20°C to 50°C (-4°F to 122°F), 80% RH or less (non-condensing)				
Operating environment	Indoors, pollution degree 2, altitude up to 2000 m (6562 ft.)				
Applicable standards	Safety: EN61010, EMC: EN61326				
Standard compliance	DIN 5032-7: 1985 class B, JIS C 1609-1: 2006 general AA class				
Dust proof and waterproof	IP40 (EN60529)				
Dimensions and mass	Approx. 78W × 170H × 39D mm (3.07" W × 6.69" H × 1.54" D)				
(including the batteries)	Approx. 310 g (10.9 oz.) (FT3424)/320 g (11.3 oz.) (FT3425)				
	Instruction Manual ×1, AA/LR6 alkaline battery ×2, Sensor cap (with				
<b>A</b>	strap) ×1, Carrying case (soft) ×1, Strap ×1, USB cable (0.9 m) ×1,				
Accessories	CD-R (USB driver, dedicated computer application software, and communications specifications) ×1, Precautions Concerning Use of				
	Equipment that Emits Radio Waves ×1 (FT3425 only)				
Interfaces	USB2.0 (FT3424/FT3425), Bluetooth® 4.0LE (only FT3425)				
Bluetooth®communication software					
Supported OS					
Supported Android devices					

#### Model: LUX METER FT3424, FT3425

Model No. (Order Code) (Note)

**EXTENSION CART Z5023** 

This cart with caster wheels can be easily

moved between measurement locations. Use with the Connection Cable L9820 to check


instrument readings from a standing posture. (The FT3425 can be paired with a smartphone,


eliminating the need for a connection cable.) Extension pole length: Approx. 0.5 m to 1.6 m

FT3424

FT3425 Built-in Bluetooth® wireless technology

#### Relative Spectral Response Characteristics in the Visible Spectrum





#### Oblique incident light characteristics

Angle	Deviation from cosine characteristics
30°	±2 %
60°	±7 %
80°	±25 %

Graph illustrates typical characteristics. Characteristics exhibited by individual products may vary slightly.

■ Data can be downloaded to tablets and smartphones using Hioki's dedicated apps available from the Google Play or App Store. (FT3425 only) Get IT ON Google Play Search for "HIOKI" and download the "GENNECT Cross" app.

\*Android, Google Play and the Google Play log oar trademarks of Google Inc.
\*Aldroid, Google Play and the Google Play log oar trademarks of Google Inc.
\*Alors is a registered trademark of Cisco Technology, Inc. and/or its affiliates in the United States and certain other countries.
\*Alphone, Plad, Irbad mini, Irbad Pro and Irbad touch are trademarks of Apple Inc.
\*Apple and the Apple log oare trademarks of Apple Inc. App Store is a service mark of Apple Inc.
\*Microsoft, Windows Wista, and Excel are either registered trademarks or trademarks of Microsoft Corporation in the United States and/or other countries.
\*Company names and Product names appearing in this catalog are trademarks or registered trademarks of various companies.
\*The Bluetooth® word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. and any use of such marks by HIOKI E.E. CORPORATION is under license.
\*For the latest information about countries and regions where wireless operation is currently supported, please visit the Hioki website.

**Options** 





Mounting method of instrument

Connection Cable L9820

Use when positioning the sensor unit and display unit separately during use (length: 2 m)



Carrying case C0202 (Soft case)

Handy for storing the instrument with the Output Cord L9094, USB cable, and Connection Cable L9820. 145W x 210H x 70D mm (5.7" W x 8.27" H x 2.76" D)



Carrying case C0201 (Semi-hard case) Stores the Output Cord L9094 and a USB cable.

137W x 193H x 69D mm (5.4" W × 7.60" H × 2.72" D)











Connect to terminal block 1.5 m(4.92 ft) length

展示機訓練等,請透過以下方式和我們聯繫,我們將真誠地爲您服務。 資料索取、産品詢問、



堉宸科技股份有限公司 02-2995-2696 www.yuctech.com.tw LINE ID: @678pknts

